
# Model KBF-S 1020 | Constant climate chambers with large temperature / humidity range

The Solid.Line KBF-S climate chamber is a custom-made solution for long-term testing and accelerated stability testing of pharmaceutical products in accordance with the ICH Q1A guideline. The corrosion-free and condensation-free interior also ensures exceptionally stable temperature and moisture levels, even when fully loaded.

#### BENEFITS

- Safe thanks to climatic homogeneity that far exceeds the accuracy required by ICH guidelines, even with a fully loaded unit.
- Reliable thanks to failsafe operation without compromise. The interior and Longlife evaporator plate are made entirely from stainless steel.
- Smart, as a wide range of accessories makes it highly compatible for adaptation to specific customer requirements.
- Economical thanks to maximum effective volume which allows for 30% larger load compared to the competition.





Model 1020

Model 1020

#### **IMPORTANT FEATURES**

- Temperature range: o °C to +70 °C
- Humidity range: 20% to 80% RH
- APT.line<sup>™</sup> preheating chamber technology
- Humidity regulation with capacitive humidity sensor and steam humidification  $\, \cdot \,$
- LCD to display temperature and humidity along with additional information and alarms
- Internal data logger, measured values can be read out in open format via USB •
- · Unit self-test for comprehensive status analysis

- Inner chamber made completely of stainless steel
- 2 stainless steel racks, incl. shelf supports
- Access port with silicone plug, 30 mm, left
- Robust twin casters with brakes, from 240 L
- Class 3.1 independent temperature safety device (DIN 12880) with visual and audible temperature alarm
- Computer interface: Ethernet
- Independent water supply via tank

#### ORDERING INFORMATION

| Interior volume<br>[L] | Power supply - unit fuse     | Plug*     | Version  | Model version   | ArtNo.    |
|------------------------|------------------------------|-----------|----------|-----------------|-----------|
| Model KBF-S 1020       |                              |           |          |                 |           |
|                        | 200230 V 1~ 50/60 Hz -16,0 A | CEE 7/7   | Standard | KBFS1020-230V   | 9020-0372 |
| 1,020                  | 200240 V 1~ 50/60 Hz -16,0 A | NEMA 6-20 | Standard | KBFS1020UL-240V | 9020-0373 |

### BINDER Best conditions for your success


# TECHNICAL DATA

| Data  |               |                                       |
|---|---------------|---------------------------------------|
| Designation   | KBFS1020-230V | KBFS1020UL-240V                       |
| Article Number  | 9020-0372     | 9020-0373                             |
| Option model  | Standard      | Standard                              |
| Performance Data Temperature                                    |               |                                       |
| Temperature range   | 070 °C        | o70 °C                                |
| Max. heat compensation at 40°C                                  | 550 W         | 550 W                                 |
| Performance Data Climate  |               |                                       |
| Temperature range with humidity                                 | 1070 °C       | 1070 °C                               |
| Humidity range  | 2080 % RH     | 2080 % RH                             |
| Temperature uniformity at 25°C and 60% RH                       | 0.3 ± K       | 0.3 ± K                               |
| Temperature uniformity at 40°C and 75% RH                       | 0.3 ± K       | 0.3 ± K                               |
| Temperature fluctuation at 25°C and 60% RH                      | 0.1 ± K       | 0.1 ± K                               |
| Temperature fluctuation at 40°C and 75% RH                      | 0.1 ± K       | 0.1 ± K                               |
| Recovery time after door was opened for 30 s at 25°C and 60% RH | 11 min        | 11 min                                |
| Recovery time after door was opened for 30 s at 40°C and 75% RH | 12 min        | 12 min                                |
| Electrical data   |               |                                       |
| Rated Voltage   | 200230 V      | 200240 V                              |
| Power frequency   | 50/60 Hz      | 50/60 Hz                              |
| Nominal power   | 2 kW          | 2 kW                                  |
| Unit fuse   | 16,0 A        | 16,0 A                                |
| Phase (Nominal voltage)   | 1~            | 1~                                    |
| Dimensions and weights  |               |                                       |
| Interior volume   | 1,020 L       | 1,020 L                               |
| Net weight of the unit (empty)                                  | 379 kg        | 379 kg                                |
| Load per rack   | 45 kg         | 45 kg                                 |
| Permitted load  | 150 kg        | 150 kg                                |
| Wall clearance back   | 100 mm        | 100 mm                                |
|   |               | · · · · · · · · · · · · · · · · · · · |


| Wall clearance sidewise                               | 300 mm   | 300 mm   |
|---|----------|----------|
| Housing dimensions not incl. fittings and connections |          |          |
| Width net   | 1,250 mm | 1,250 mm |
| Height net  | 1,925 mm | 1,925 mm |
| Depth net   | 1,145 mm | 1,145 mm |
| Internal Dimensions                                   |          |          |
| Interior width  | 976 mm   | 976 mm   |
| Interior height                                       | 1,250 mm | 1,250 mm |
| Interior depth  | 836 mm   | 836 mm   |
| Unit doors  | 2        | 2        |
| Environment-specific data                             |          |          |
| Sound-pressure level                                  | 56 dB(A) | 56 dB(A) |
| Energy consumption at 40°C and 75% RH                 | 550 Wh/h | 550 Wh/h |
| Fixtures  |          |          |
| Number of shelves (std./max.)                         | 2/15     | 2/15     |

All technical data is specified for unloaded units with standard equipment at an ambient temperature of  $\pm 22 \,^{\circ}C \pm 3 \,^{\circ}C$  and a power supply voltage fluctuation of  $\pm 10$  %. The temperature data is determined in accordance to BINDER factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. Technical data refers to 100 % fan speed. All indications are average values, typical for units produced in series. We reserve the right to change technical specifications at any time.

#### DIMENSIONS Incl. fittings and connections [mm]



#### BINDER Best conditions for your success



# OPTIONS

| Designation  | Description  | KBF-S 1020 | *  | ArtNo.    |
|--|--|------------|----|-----------|
|  | left   |            |    |           |
|  | 30 mm  | •          | 01 | 8012-2072 |
|  | 50 mm  | •          | 01 | 8012-2079 |
|  | 100 mm   | •          | 01 | 8012-2077 |
|  | right  |            |    |           |
| Access port with silicone plug                       | 30 mm  | •          | 01 | 8012-2071 |
| Access port with shicone plug                        | 50 mm  | •          | 01 | 8012-2078 |
|  | 100 mm   | •          | 01 | 8012-2076 |
|  | top  |            |    |           |
|  | 30 mm  | •          | 01 | 8012-2073 |
|  | 50 mm  | •          | 01 | 8012-2074 |
|  | 100 mm   | •          | 01 | 8012-2075 |
| Calibration certificate,<br>expanded                 | for temperature and humidity; for extending the measurement in center of chamber to include another test value | •          | -  | 8012-2066 |
| Calibration certificate,<br>temperature              | temperature measurement incl. certificate, 9 measuring points at specified temperature                         | •          | -  | 8012-2067 |
|  | temperature measurement incl. certificate, 15- 18 measuring points at specified temperature                    | •          | -  | 8012-2068 |
|  | temperature measurement incl. certificate and 27 measuring points at specified temperature                     | •          | -  | 8012-2069 |
| Calibration certificate,<br>temperature and humidity | Measurement in center of chamber at 25°C / 60% RH or at specified test values                                  | •          | -  | 8012-2065 |

Data sheet Model KBF-S 1020

# BINDER

| Best | conditions | for | your | success |  |
|------|------------|-----|------|---------|--|
|      |            |     |      |         |  |

| Designation | Description   | KBF-S 1020 | * | ArtNo.    |
|-------------|---|------------|---|-----------|
|             | temperature (according to DIN12880) and humidity measurement incl. certificate, 27<br>temperature measuring points and 1 humidity measuring point, at 25 °C / 60 % RH or at<br>specified values | •          | - | 8012-2070 |
| Door lock   | lockable door handle  | •          | - | 8012-2087 |

## ACCESSORIES

| Designation                              | Description  | KBF-S 1020 | * | ArtNo.    |
|--|--|------------|---|-----------|
|  | for simple logging and documentation requirements with up to 5 networked units.  |            |   |           |
| APT-COM™ 4 BASIC-Edition                 | version 4, BASIC edition   | •          | _ | 9053-0039 |
| APT-COM™ 4 GLP-Edition                   | for working under GLP-compliant conditions. Measured values are documented in a tamper-<br>proof way in line with the requirements of FDA Regulation 21 CFR 11.  |            |   |           |
|  | version 4, GLP edition   | •          | _ | 9053-0042 |
| APT-COM™ 4 PROFESSIONAL-<br>Edition      | convenient unit and user management built on the BASIC edition. Suitable for networking up to 100 units.   |            |   |           |
| Edition                                  | version 4, PROFESSIONAL edition  | •          | _ | 9053-0040 |
| BINDER PURE AQUA SERVICE                 | System for preparation or complete desalination of tap water, complete set containing PURE AQUA 300 single-use cartridge, measuring device, and all necessary connecting parts   | •          | - | 8012-0759 |
| BINDER PURE AQUA SERVICE,<br>accessories | Single-use, replacement cartridge for BINDER PURE AQUA System  | •          | - | 6011-0165 |
| Convenient water supply                  | Flexible positioning of the water container with magnetic holder, hose adapter for water container with practical quick coupling, and water stop function  |            |   |           |
|  | External, with magnetic holder   | •          | - | 8012-1847 |
|  | Basic set consisting of 2 pieces, attachment material, control unit for max. 4 light strips, 100-240 V, 50/60 Hz   |            |   |           |
|  | Basic set 300, length 30 cm  | •          | - | 8012-1107 |
|  | Basic set 500, length 50 cm  | •          | - | 8012-1108 |
|  | Basic set 900, length 90 cm  | •          |   | 8012-1109 |
| LED light bars                           | Expansion set consisting of 2 pieces, attachment material: clips. For expanding the basic set of light bars  |            |   |           |
|  | Expansion set 300, length 30 cm  | •          | - | 8012-1716 |
|  | Expansion set 500, length 50 cm  | •          |   | 8012-1717 |
|  | Expansion set 900, length 90 cm  | •          | - | 8012-1718 |
| Qualification documents                  | IQ/OQ/PQ documents – supporting documents for validation performed by customers,<br>according to customer requirements, PQ section added to qualification folder IQ/OQ;<br>parameters: temperature and humidity values |            |   |           |

Data sheet Model KBF-S 1020

#### BINDER Best conditions for your success

| Designation         | Description   | KBF-S 1020 | * | ArtNo.    |
|---------------------|---|------------|---|-----------|
|                     | Digital in PDF format   | •          |   | 7057-0006 |
|                     | Hard copy inside folder   | •          | - | 7007-0006 |
|                     | IQ/OQ documents – supporting documents for validation performed by customers, consisting of: IQ/OQ checklists incl. calibration guide and comprehensive unit documentation; parameters: temperature and humidity values |            |   |           |
|                     | Digital in PDF format   | •          |   | 7057-0002 |
|                     | Hard copy inside folder   | •          | - | 7007-0002 |
| Rack                | stainless steel   | •          | - | 8012-2052 |
| Rack, reinforced    | stainless steel, with fasteners (1 set of 4)  | •          | - | 8012-0968 |
| Rack accessories    | fasteners (1 set of 4) for additional security of racks   | •          |   | 8012-2280 |
| Shelf, perforated   | Stainless steel   | •          |   | 8012-2254 |
| H-neutral detergent | concentrated, for gentle remove of residual contaminants; 1 kg  | •          | - | 8012-2250 |

# SERVICES

| Designation  | Description  | * | ArtNo.    |
|--|--|---|-----------|
| Calibration services   |  |   |           |
| Townser two and humidity calibration                           | Expansion –Temperature and humidity calibration with 1 measuring point in center of chamber with 1 specified pair of values, including certificate   | - | DL30-0302 |
| Temperature and humidity calibration                           | Temperature and humidity calibration with 1 measuring point in center of chamber with 1 specified pair of values, including certificate  | _ | DL30-0301 |
| Temperature and humidity measurement, 9-1<br>measuring points  | Temperature measurement with 9 temperature measuring points and 1 humidity measuring point in center of chamber with a pair of values specified by the user, including certificate   | _ | DL30-0309 |
| Temperature and humidity measurement, 18-1<br>measuring points | Temperature measurement with 18 temperature measuring points and 1 humidity measuring point in center of chamber with a pair of values specified by the user, including certificate  | _ | DL30-0318 |
| Temperature and humidity measurement, 27-1 measuring points    | Temperature measurement with 27 temperature measuring points and 1 humidity measuring point in center of chamber with a pair of values specified by the user, including certificate  | _ | DL30-0327 |
| Temperature and humidity measurement according to DIN12880     | Temperature measurement in accordance with DIN 12880 with 27 temperature measuring points and 1 humidity measuring point in center of chamber with a pair of values specified by the user, including certificate                 | _ | DL30-0427 |
| Installation services  |  |   |           |
| Unit commissioning   | Connect the unit to the customer-side connections (electricity, water, wastewater, gas), basic functions check, brief operating instructions. (excl.: unpacking, setup, controller instructions, programming, installation work) | _ | DL10-0300 |
| Unit instructions  | Instruction regarding operating principle and basic functions of the unit, operation of the control electronics including programming  | _ | DL10-0700 |

Data sheet Model KBF-S 1020

#### BINDER Best conditions for your success

| ArtNo.    |
|-----------|
| DL20-0710 |
| DL20-0920 |
| DL20-0820 |
|           |
| DL20-0400 |
|           |
| DL42-0300 |
| DL44-0500 |
|           |
| DL50-0030 |
|           |

# NOTES

01 Condensation may occur in the area around the access port. Access ports may be placed in custom locations for an additional charge.

# **BINDER GmbH**

Tuttlingen, Germany TEL +49 7462 2005 0 FAX +49 7462 2005 100 info@binder-world.com www.binder-world.com

### BINDER Asia Pacific (Hong Kong) Ltd.

Kowloon, Hong Kong, P.R. China TEL +852 39070500 FAX +852 39070507 asia@binder-world.com www.binder-world.com

#### **BINDER Inc.**

Bohemia, NY, USA TEL +1 631 224 4340 FAX +1 631 224 4354 usa@binder-world.com www.binder-world.us

# **BINDER Environmental Testing**

Equipment (Shanghai) Co., Ltd.

Shanghai, P.R. China TEL +86 21 685 808 25 FAX +86 21 685 808 29 china@binder-world.com www.binder-world.com